



Technology for Precision Measurements



PHASE DYNAMICS LOW RANGE INSERTION MODEL NUMBERS

LI A A B C D E F G H I J K

Field AA designates process connection:

1R	- ANSI Class 150 RF Flange	1J	-RTJ	1F	-Flat Face
3R	- ANSI Class 300 RF Flange	3J	- RTJ	3F	-Flat Face
6R	- ANSI Class 600 RF Flange	6J	- RTJ	6F	-Flat Face
9R	- ANSI Class 900 RF Flange	9J	- RTJ	9F	-Flat Face

Field B designates measurement section diameter:

3 - 3 inch (with RTD Assembly)

Field C designates measurement section material:

0	- 316 / 316L Stainless	2	- Monel
1	- Duplex 2205	3	- Hastelloy

Field D designates measurement section configuration:

1	- 160°F (71°C) Maximum 6" Active Region Length, 11.9 inches overall.
2	- 220°F (104°C) Maximum 6" Active Region Length, 11.9 inches overall.
4	- 400°F (204°C) Maximum 6" Active Region Length, 11.9 inches overall.
6	- 600°F (315°C) Maximum 6" Active Region Length, 11.9 inches overall.

Field E designates measurement range:

0	- 0 to 4% Water in Oil (standard)
1	- 0 to 10% Water in Oil
2	- 0 to 20% Water in Oil

Field F designates electronics enclosure type:

4	- NEMA 4X (only certification available is CSA)	X	- No Electronics Enclosure
E	- Explosion Proof		

Field G designates electronics system:

X	- Standard Electronics and Standard Oscillator Module
S	- Stand-alone Standard Oscillator Module

Field H designates system cable length:

0	- 30 feet (standard length)	5	- 30 feet Low Temp
1	- 50 feet	6	- 50 feet Low Temp
2	- 100 feet	7	- 100 feet Low Temp
3	- 150 feet (maximum length)	8	- 150 feet Low Temp (maximum length)
X	- No Cable		

Field I designates certification requirements:

1	- PDI – ¾"NPT Conduit Entry	5	- PDI – M20 Conduit Entry
2	- Factory Mutual – ¾"NPT Conduit Entry	6	- CE EEx II 2 G/D ATEX - ¾"NPT Conduit Entry
3	- CSA – ¾"NPT Conduit Entry		
4	- CE EEx II 2 G/D ATEX – M20 Conduit Entry		

Field J designates communication protocols:

0	- RS-422 ASCII	2	- MODBUS® RTU RS-485 (Standard)
1	- RS-485 ASCII (Multi Drop)	3	- HART®

Field K designates power requirements:

blank	- 120 VAC 60 Hz (standard)	D	- 120 VAC 60Hz with Enclosure Heater
A	- 230 VAC 50 Hz	E	- 230 VAC 50 Hz with Enclosure Heater
B	- 24 VDC	F	- 24 VDC with Enclosure Heater
C	- 120 VAC 50 Hz	G	- 120 VAC 50 Hz with Enclosure Heater