



Technology for Precision Measurements



PHASE DYNAMICS FULL RANGE MODEL NUMBERS

FL AA B C D E F G H I J K "L" CONFIGURATION
FU AA B C D E F G H I J K "U" CONFIGURATION
FZ AA B C D E F G H I J K "Z" CONFIGURATION

Field AA designates process connection:

1R	- ANSI Class 150 RF Flange	1J	- RTJ	1F	- Flat Face
3R	- ANSI Class 300 RF Flange	3J	- RTJ	3F	- Flat Face
6R	- ANSI Class 600 RF Flange	6J	- RTJ	6F	- Flat Face
9R	- ANSI Class 900 RF Flange	9J	- RTJ	9F	- Flat Face
5R	- ANSI Class 1500 RF Flange	5J	- RTJ	5F	- Flat Face

Field B designates measurement section diameter:

2	- 2 inch
3	- 3 inch
4	- 4 inch

Field C designates measurement section material:

0	- 316 / 316L Stainless	2	- Monel
1	- Duplex 2205	3	- Hastelloy

Field D designates measurement section temperature range:

1	- 160° F (71°C) Maximum	4	- 400° F (204°C) Maximum
2	- 220° F (104°C) Maximum	6	- 600° F (315°C) Maximum

Field E designates salinity range (See Salinity Matrix for salinity vs. temperature):

0	- 0.5 to 8.0% Salinity	5	- 15.0 to 28.0% Salinity
1	- 0.1 to 0.5% Salinity	6	-
2	- 0.1 to 3.0% Salinity	7	-
3	- 0.1 to 8.0% Salinity	8	-
4	- 8.0 to 25.0% Salinity	9	- Special-Consult Factory

Field F designates electronics enclosure type:

4	- NEMA 4X (only certification available is CSA)	X	- No Electronics Enclosure
E	- Explosion Proof		

Field G designates electronics system:

X	- Standard Electronics and Standard Oscillator Module
S	- Stand-alone Standard Oscillator Module

Field H designates system cable length:

0	- 30 feet (standard length)	5	- 30 feet Low Temp
1	- 50 feet	6	- 50 feet Low Temp
2	- 100 feet	7	- 100 feet Low Temp
3	- 150 feet (maximum length)	8	- 150 feet Low Temp (maximum length)
X	- No Cable		

Field I designates certification requirements:

1	- PDI – ¾" NPT Conduit Entry	5	- PDI – M20 Conduit Entry
2	- Factory Mutual – ¾" NPT Conduit Entry	6	- CE EEx II 2 G/D ATEX - ¾" NPT Conduit Entry
3	- CSA – ¾" NPT Conduit Entry		
4	- CE EEx II 2 G/D ATEX – M20 Conduit Entry		

Field J designates communication protocols:

0	- RS-422 ASCII	2	- MODBUS® RTU RS-485 (Standard)
1	- RS-485 ASCII (Multi Drop)	3	- HART®

Field K designates power requirements:

blank	- 120 VAC 60 Hz (standard)	D	- 120 VAC 60Hz with Enclosure Heater
A	- 230 VAC 50 Hz	E	- 230 VAC 50 Hz with Enclosure Heater
B	- 24 VDC	F	- 24 VDC with Enclosure Heater
C	- 120 VAC 50 Hz	G	- 120 VAC 50 Hz with Enclosure Heater